



→ IDENTIFICATION

endo revision

→ MORPHOLOGY

angled insert with 0.6 mm Ø, working length up to 10 mm, the last 5 mm D30 diamond coated

→ SURFACE

titanium nitride coat

→ TREATMENT

presentation of calcified root canals entrances, removal of root elements and fractured inserts in the coronal third of the root canal

→ POWER

0 – 50 %

