

## → IDENTIFICATION

endo revision

## → MORPHOLOGY

angled insert with 0.6 mm Ø, working length up to 20 mm, the last 5 mm D30 diamond coated

## → SURFACE

titanium nitride coat

## → TREATMENT

presentation of calcified root canals entrances,  
removal of root elements and fractured inserts  
in the coronal third of the root canal

## → POWER

0 – 50 %

