



→ IDENTIFICATION

endo revision

→ MORPHOLOGY

angled shaft with small, spherical ( $\varnothing$  1.7 mm)  
and D30 diamond coated tip

→ SURFACE

titanium nitride coat

→ TREATMENT

exposure of root canal pins

→ POWER

0 – 50 %