



→ IDENTIFICATION

endo revision

→ MORPHOLOGY

angled insert with 0.5 mm  $\emptyset$ , working length up to 24 mm

→ SURFACE

titanium nitride coat

→ TREATMENT

removal of fractured inserts in the medial and apical third of the root canal

→ POWER

0 – 50 %